



Software and Firmware Update

Release of Noxturnal 7.3

At Nox Medical, we are committed to continuous innovation and system reliability. With the release of **Noxturnal 7.3**, users benefit from important stability and performance improvements, including the correction of previously observed asynchronous flow behavior, along with targeted workflow and reporting enhancements.

Noxturnal 7.3 contains updates of the following:

- Noxturnal Software
- Nox A1s Firmware
- Nox T3s Firmware

Downloading the Software

To download the Noxturnal 7.3 software and firmware updates, please visit our website.

Within our Release Bulletin Hub, you can access a comprehensive document detailing the new features, enhancements, fixes, and compatibility information included in the Noxturnal 7.3 release.

<https://noxmedical.com/downloads>

<https://noxmedical.com/release-bulletin-hub/>

Main updates include:

- » Important stability improvements and performance enhancements, including correction of previously observed asynchronous flow behavior.
- » Improvements to recording library indexing and device communication reliability.
- » New feature: Hypoxic burden reporting for REM/NREM and Supine/Non-Supine conditions.
- » New feature: Automatic association of arousals and PLMs.
- » New feature: Tech notes remain attached to recordings across scoring sessions.
- » Improvements to REI calculations including 3% and 4% desaturation metrics.
- » Enhanced signal visualization with improved axis line configurations and customizable spectrogram settings.

DISCLAIMER: To take advantage of these new features and improvements, update to **Noxturnal 7.3** and ensure your **Nox T3s** and **Nox A1s** devices are updated to the latest firmware. Please refer to the Release Bulletin for compatibility and upgrade guidance.